

*SOURCE: Miss. Code Ann. §41-59-5*

Rule 1.5.7 Standard mandatory miscellaneous equipment: Unless otherwise precluded elsewhere in this specification, each ambulance shall be equipped with, but not limited to, the following:

1. Fire extinguisher: one, ABC dry chemical, multi-purpose (Halon, CO<sub>2</sub>) minimum 5-pound unit in a quick-release bracket mounted in the patient compartment.
2. Hazardous Material Reference Guide.
3. Reflective Safety wear for each crewmember (must meet or exceed ANSI/ISEA performance class II or III).

*SOURCE: Miss. Code Ann. §41-59-5*

Rule 1.5.8 Medical, surgical, and bio-medical equipment for all levels: When specified, the ambulance shall be equipped with, but not limited to, the following: #24068

1. One stretcher for primary patient as specified in current Federal Specifications for ambulances, dimensions as per KKK-A-1822.
2. 3 strap type restraining devices (chest, hip, knee, and shoulder) attached to stretcher. Straps shall not be less than two inches wide, nylon, and consist of quick release buckles.
3. Portable and fixed oxygen equipment with variable flow regulator capable of delivering 15 lpm in calibrated increments. Cylinder must contain 300 psi of medical grade O<sub>2</sub> at a minimum.
4. Three oxygen masks, adult. (Non-rebreathing face mask)
5. One oxygen mask, child. (Non-rebreathing face mask)
6. One oxygen mask, infant.
7. Three oxygen bi-pronged nasal cannulas.
8. One oxygen bi-pronged nasal cannula - pediatric.
9. One mouth-to-mask artificial ventilation device with supplemental oxygen inlet port with one-way valve, i.e., "pocket mask", etc.
10. Bag Valve Mask (manual resuscitator) hand operated, self reexpanding bag, adult (>1000 ml), without pop-off valve, with oxygen reservoir capable of delivering 80-100 percent oxygen.